

## Research Kit Specifics

Catalog Number	Salivary Assay Kit	Serum-Saliva Correlation	Calibrator Range	Sensitivity	Salivary Test Volume	Incubation Time	Number of Samples (singlets)
1-1902	α Amylase	N/A	N/A	N/A	8 μL of X200 dilution	3 minutes	94
1-2902	Androstenedione	0.77	10 - 2430 pg/mL	5 pg/mL	50 μL	2.5 hours	76
1-1302	Blood Contamination	N/A	0.08 - 6.6 mg/dL	0.08 mg/dL	20 μL	1.0 hour	80
1-3002	ER HS Cortisol	0.91	0.012 - 3.000 μg/dL	<0.003 μg/dL	25 μL	1.5 hours	76
1-2002	Cotinine	N/A	0.8 - 200 ng/mL	0.05 ng/mL	20 μL	2.0 hours	76
1-3302	C-Reactive Protein	N/A	93.75 - 3000 pg/mL	10 pg/mL	50 μL of X10 dilution	2.5 hours	78
1-1202	DHEA	0.86	10.2 - 1000 pg/mL	5 pg/mL	50 μL	3.5 hours	76
1-1252	DHEA-S	N/A	188.9 - 15,300 pg/mL	43 pg/mL	100 μL	1.5 hours	78
1-3702	HS Estradiol	0.80	1.0 - 32.0 pg/mL	<1 pg/mL	100 μL	2.5 hours	76
1-1802	Estriol	0.87	20 - 4860 pg/mL	16 pg/mL	25 μL	2.5 hours	76
1-1802	HS Estriol	0.87	5 - 1215 pg/mL	1 pg/mL	100 μL of X2 dilution	21 - 25 hours	76
1-3202	Estrone	N/A	3.1 - 300 pg/mL	<1 pg/mL	100 μL	3 hours	76
1-1502	Progesterone	0.87	10 - 2430 pg/mL	5 pg/mL	50 μL	1.5 hours	76
1-2602	17-α-Hydroxy-Progesterone	0.87	5.1 - 500 pg/mL	3 pg/mL	50 μL	2.5 hours	76
1-1602	SlgA	NA	2.5 - 600 μg/mL	2.5 μg/mL	25 μL of X2000 dilution	3.75 hours	76
1-2402	ER Testosterone	0.96	6.1 - 600 pg/mL	1 pg/mL	25 μL	1.5 hours	76

All Salimetrics kits include controls.